

## **Appendix D. Calibration Certificates**

## EQUIPMENT CALIBRATION RECORD

Type : Laser Dust Monitor  
 Manufacturer / Brand : SIBATA  
 Model No.: LD-3B  
 Equipment No.: LD-3B-002  
 Serial No.: 974350  
 Sensitivity Adjustment Scale Setting : 622 CPM

### Standard Equipment

Equipment : MFC High Volume Air Sampler  
 Venue : Tung Chung Pier  
 Model No.: TE-5170 Total Suspended Particulate  
 Serial No.: S/N3641  
 Previous Calibration Date: 12/Jul/2018

### Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration) : 624 CPM  
 Sensitivity Adjustment Scale Setting (After Calibration) : 624 CPM


Hour	Date (dd-mm-yyyy)	Time		Ambient Condition		Concentration ( $\mu\text{g}/\text{m}^3$ ) Y-axis	Total Count	Count/Minute X-axis
				Temp ( $^{\circ}\text{C}$ )	R.H. (%)			
1	24-Aug-18	09:07	09:37	29.4	80%	78.1	1513	50.4
2	24-Aug-18	10:00	11:00	30.5	76%	132.7	7857	131.0
3	24-Aug-18	11:12	12:42	30.6	76%	147.4	12486	138.7
4	24-Aug-18	13:21	15:21	31.0	71%	211.4	27133	226.1

Be Linear Regression of Y or X

Slope (K-factor): 0.9945

Correlation coefficient (R): 0.9788

Remark: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_



Recorded by: Shing Mak

Signature: *Shing*

Date: 07/Sep/2018

Checked by: Eva Keung

Signature: *Eva*

Date: 07/Sep/2018

## EQUIPMENT CALIBRATION RECORD

Type : Laser Dust Monitor  
 Manufacturer / Brand : SIBATA  
 Model No.: LD-3B  
 Equipment No.: LD-3B-003  
 Serial No.: 276018  
 Sensitivity Adjustment Scale Setting : 799 CPM

### Standard Equipment

Equipment : MFC High Volume Air Sampler  
 Venue : Tung Chung Pier  
 Model No.: TE-5170 Total Suspended Particulate  
 Serial No.: S/N3641  
 Previous Calibration Date: 12/Jul/2018

### Calibration Result

Sensitivity Adjustment Scale Setting (Before Calibration) : 701 CPM  
 Sensitivity Adjustment Scale Setting (After Calibration) : 701 CPM


Hour	Date (dd-mm-yyyy)	Time		Ambient Condition		Concentration (ug/m <sup>3</sup> ) Y-axis	Total Count	Count/Minute X-axis
				Temp (°C)	R.H. (%)			
1	24-Aug-18	09:07	09:37	29.4	80%	78.1	1432	47.7
2	24-Aug-18	10:00	11:00	30.5	76%	132.7	6251	104.2
3	24-Aug-18	11:12	12:42	30.6	76%	147.4	10055	111.7
4	24-Aug-18	13:21	15:21	31.0	71%	211.4	20762	173.0

Be Linear Regression of Y or X

Slope (K-factor): 1.2709

Correlation coefficient (R): 0.9919

Remark: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_



Recorded by: Icy Chan

Signature: *Icy Chan*

Date: 07/Sep/2018

Checked by: Eva Keung

Signature: *Eva Keung*

Date: 07/Sep/2018

ENVIROTECH SERVICES CO.

**High-Volume TSP Sampler**  
**5-Point Calibration Record**

Location : AMS2(Tung Chung Development Pier)  
Calibrated by : P. F. Yeung  
Date : 12/07/2018

**Sampler**

Model : TE-5170  
Serial Number : S/N3641

**Calibration Orifice and Standard Calibration Relationship**

Serial Number : 2454  
Service Date : 19 Mar 2018  
Slope (m) : 2.05242  
Intercept (b) : -0.01383  
Correlation Coefficient(r) : 0.99994

**Standard Condition**

Pstd (hpa) : 1013  
Tstd (K) : 298.18

**Calibration Condition**

Pa (hpa) : 1003  
Ta(K) : 304

Resistance Plate	dH [green liquid] (inch water)	Z	X=Qstd (cubic meter/min)	IC	Y
1   18 holes	12.8	3.525	1.724	56	55.17
2   13 holes	10.2	3.146	1.540	50	49.26
3   10 holes	7.8	2.751	1.347	46	45.32
4   7 holes	4.8	2.158	1.058	38	37.44
5   5 holes	3.2	1.762	0.865	30	29.56

Notes:  $Z = \sqrt{dH(Pa/Pstd)(Tstd/Ta)}$ ,  $X = Z/m - b$ ,  $Y(\text{Corrected Flow}) = IC * \{\sqrt{Pa/Pstd}(Tstd/Ta)\}$

**Sampler Calibration Relationship**

Slope(m): 28.663      Intercept(b): 5.886      Correlation Coefficient(r): 0.9950

Checked by: Magnum Fan

Date: 16/07/2018



專業化驗有限公司

**QUALITY PRO TEST-CONSULT LIMITED**

Unit 10, 14/F, Wah Wai Centre, 38-40 Au Pui Wan St., Fotan, Hong Kong

Email: info@qualityprotest.com; Website: www.qualityprotest.com

Tel: (852) 3956 8717; Fax: (852) 3956 3928

## CALIBRATION REPORT

Test Report No. : AG090183  
Date of Issue : 19 September, 2018  
Page No. : 1 of 2

### PART A – CUSTOMER INFORMATION

Enovative Environmental Service Ltd.  
Rm 811, Hin Pui House,  
Hin Keng Estate, Tai Wai  
New Territories, Hong Kong  
Attn: Mr. Thomas Wong

### PART B – SAMPLE INFORMATION

Description of Samples : Titrette® bottle-top burette, 50mL  
Brand Name : BRAND  
Model Number : 6761161  
Manufacturer number : 4760161  
Serial Number : 10N64701  
Date of Received : Sep 15, 2018  
Date of Calibration : Sep 18, 2018  
Date of Next Calibration<sup>(a)</sup> : Dec 18, 2018

### PART C – CALIBRATION REQUESTED

<u>Parameter</u>	<u>Reference Method</u>
Accuracy Test	In-house Method (Gravimetric Method)


~ Continued On Next Page ~

Remark(s): -

<sup>(a)</sup> The "Date of Next Calibration" is recommended according to best practice principals as practiced by QPT or quoted from relevant international standards.

APPROVED SIGNATORY :

FUNG Yuen-ching Aries  
Laboratory Manager



## CALIBRATION REPORT

Test Report No. : AG090183  
 Date of Issue : 19 September, 2018  
 Page No. : 2 of 2

### PART D – RESULT<sup>(b),(c)</sup>

Water temperature: 22.4 °C

Relative humidity: 51%

z-Factor: 1.0034

Trial	Nominal volume (mL) at interval				
	3	3	3	3	3
	Range: (1-4)	Range: (16-19)	Range: (23-26)	Range: (34-37)	Range: (42-45)
1	2.9816	2.9866	2.9815	2.9821	2.9814
2	2.9850	2.9852	2.9814	2.9833	2.9881
3	2.9912	2.9830	2.9885	2.9867	2.9889
4	2.9855	2.9867	2.9877	2.9914	2.9808
5	2.9882	2.9843	2.9793	2.9921	2.9883
6	2.9895	2.9908	2.9850	2.9878	2.9878
7	2.9941	2.9909	2.9873	2.9890	2.9822
8	2.9921	2.9891	2.9852	2.9862	2.9892
9	2.9888	2.9902	2.9878	2.9883	2.9858
10	2.9931	2.9921	2.9855	2.9938	2.9818
Average	2.9889	2.9879	2.9849	2.9881	2.9854
Standard deviation	0.0040	0.0031	0.0032	0.0037	0.0035
Calculated volume (mL)	2.9991	2.9980	2.9951	2.9982	2.9956
Error (%)	-0.0309	-0.0650	-0.1644	-0.0590	-0.1473
RSD (%)	0.1320	0.1050	0.1059	0.1243	0.1160

### Acceptance Criteria<sup>(d)</sup>

Accuracy (%Error)	< ±1%	< ±1%	< ±1%	< ±1%	< ±1%
Precision (%RSD)	< 1%	< 1%	< 1%	< 1%	< 1%

~ END OF REPORT ~

Remark(s): -

<sup>(b)</sup> The results relate only to the tested sample as received

<sup>(c)</sup> The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

<sup>(d)</sup> The "acceptance criteria" is applicable for similar equipment used by QPT or quoted from relevant international standards.