

## **Appendix E. Calibration Certificates**

# Manufacturer Calibration Certificate

---

The following instrument has been tested and calibrated to the manufacturer specifications.  
The calibration is traceable in accordance with ISO/IEC 17025 covering all instrument functions.

- Device Type: **XL2 Audio and Acoustic Analyzer**
- Serial Number: **A2A-14829-E0**

- Certificate Issued: **28 August 2018**
- Certificate Number: **43340-A2A-14829-E0**
- Results: **PASSED**  
(for detailed report see next page)

---

Tested by: **M. Frick**

Signature:

Stamp:



**NTi Audio AG**  
Im alten Riet 102  
LI-9494 Schaan  
[www.nti-audio.com](http://www.nti-audio.com)

Calibration of: XL2 Audio and Acoustic Analyzer  
 Serial Number: A2A-14829-E0  
 Date: 28 August 2018

• Detailed Calibration Test Results:

	reference	actual	unit	actual error	XL2 tolerance	calibration uncertainty <sup>2</sup>
RMS Level @ 1kHz, XLR Input	0.1	<b>0.100</b>	V	≤0.1%	±0.5%	±0.10%
	1	<b>1.000</b>	V	≤0.1%	±0.5%	±0.09%
	10	<b>9.988</b>	V	-0.1%	±0.5%	±0.09%
Flatness, XLR Input <sup>1</sup>	20 Hz	<b>0.996</b>	V	-0.4%	±1.1%	±0.09%
	20 kHz	<b>1.005</b>	V	0.5%	±1.1%	±0.09%
Frequency	1000	<b>999.99</b>	Hz		±0.003%	±0.01%
Residual Noise	XLR	<b>&lt; 2 uV</b>			<2 uV	±0.50%
THD+N @ 0 dBu, 1 kHz, XLR Input		<b>-98.5</b>	dB		typ. -100 dB	±0.50%

- Test Conditions: Temperature: **25.7** °C  
 Relative Humidity: **55.2** %

• Calibration Equipment Used:

- Agilent Multimeter, Typ 34401A, Serial No. MY 5300 4607  
 Last calibration: 15.08.2018, Next calibration: 15.08.2019  
 Calibrated by ELCAL to the national standards maintained at Swiss Federal Office of Metrology. SCS 0002

- FX100 Audio Analyzer, Serial No. 10408  
 Last Calibration: 27.04.2018, Next Calibration: 27.04.2019  
 Manufacturer calibration based on Agilent 34410, Serial No. MY47014254,  
 Last Calibration: 11.05.2018, Next Calibration: 11.05.2019  
 which is calibrated by ELCAL to national standards maintained at Swiss Federal Office of Metrology. SCS 002

<sup>1</sup> The specified tolerance +/-0.1 dB @ 1V = +/- 1.1%

<sup>2</sup> The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with the regulations of the GUM.



專業化驗有限公司

QUALITY PRO TEST-CONSULT LIMITED

Unit 10, 14/F, Wah Wai Centre, 38-40 Au Pui Wan St., Fotan, Hong Kong

Email: info@qualityprotest.com; Website: www.qualityprotest.com

Tel: (852) 3956 8717; Fax: (852) 3956 3928

## CALIBRATION REPORT

Test Report No. : AH110159  
Date of Issue : 04 December 2018  
Page No. : 1 of 2

### PART A – CUSTOMER INFORMATION

Enovative Environmental Service Ltd.  
Flat 2207, Yu Fun House,  
Yu Chui Court, Shatin,  
New Territories, Hong Kong  
Attn: Mr. Thomas Wong

### PART B – SAMPLE INFORMATION

Description of Samples : Titrette® bottle-top burette, 50mL  
Brand Name : BRAND  
Model Number : 6761161  
Serial Number : 10N60623  
Date of Received : Nov 28, 2018  
Date of Calibration : Nov 28, 2018  
Date of Next Calibration<sup>(a)</sup> : Feb 28, 2019

### PART C – CALIBRATION REQUESTED


<u>Parameter</u>	<u>Reference Method</u>
Accuracy Test	In-house Method (Gravimetric Method)

~ Continued On Next Page ~

Remark(s): -

<sup>(a)</sup> The "Date of Next Calibration" is recommended according to best practice principals as practiced by QPT or quoted from relevant international standards.

APPROVED SIGNATORY: \_\_\_\_\_

  
LAM Ho-ye, Emma  
Assistant Laboratory Manager



專業化驗有限公司  
**QUALITY PRO TEST-CONSULT LIMITED**

Unit 10, 14/F, Wah Wai Centre, 38-40 Au Pui Wan St., Fotan, Hong Kong  
 Email: info@qualityprotest.com; Website: www.qualityprotest.com  
 Tel: (852) 3956 8717; Fax: (852) 3956 3928

## CALIBRATION REPORT

Test Report No. : AH110159  
 Date of Issue : 04 December 2018  
 Page No. : 2 of 2

### PART D – RESULT<sup>(b),(c)</sup>

Water temperature: 25.2 °C

Relative humidity: 63%

z-Factor: 1.0039

Trial	Nominal volume (mL) at interval				
	3	3	3	3	3
	Range: (1-4)	Range: (16-19)	Range: (23-26)	Range: (34-37)	Range: (42-45)
1	2.9754	2.9751	2.9755	2.9754	2.9749
2	2.9788	2.9770	2.9666	2.9750	2.9786
3	2.9781	2.9772	2.9775	2.9771	2.9735
4	2.9740	2.9734	2.9751	2.9826	2.9737
5	2.9750	2.9764	2.9757	2.9777	2.9784
6	2.9756	2.9781	2.9818	2.9789	2.9718
7	2.9781	2.9774	2.9772	2.9761	2.9781
8	2.9756	2.9775	2.9791	2.9779	2.9795
9	2.9719	2.9810	2.9845	2.9831	2.9790
10	2.9797	2.9794	2.9835	2.9806	2.9801
Average (g)	2.9762	2.9773	2.9777	2.9784	2.9768
Standard deviation	0.0024	0.0021	0.0051	0.0028	0.0030
Calculated volume (mL)	2.9878	2.9889	2.9893	2.9901	2.9884
Error (%)	-0.4058	-0.3713	-0.3579	-0.3315	-0.3877
RSD (%)	0.0807	0.0703	0.1719	0.0953	0.0995

### Acceptance Criteria<sup>(d)</sup>

Accuracy (%Error)	< ±1%	< ±1%	< ±1%	< ±1%	< ±1%
Precision (%RSD)	< 1%	< 1%	< 1%	< 1%	< 1%

~ END OF REPORT ~

**Remark(s): -**

<sup>(b)</sup> The results relate only to the tested sample as received

<sup>(c)</sup> The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

<sup>(d)</sup> The "acceptance criteria" is applicable for similar equipment used by QPT or quoted from relevant international standards.